

Ready Snap®

Revision 02/07/13
480904 Ready Snap® 4

4

Calibration Verification Solution Instruction Manual

CAN BE USED WITH THE 638NM EXACT® MICRO
PHOTOMETERS* TO VERIFY INSTRUMENT OPERATION

*This includes models eXact® Micro 8, eXact® Micro 20, and eXact® Eco-Check.



Industrial Test Systems, Inc.

Innovators of Water Quality Testing

Manufactured By: Industrial Test Systems, Inc.
1875 Langston Street, Rock Hill, SC 29730 USA
Phone: 1-800-861-9712 - *INSIDE THE U.S.*
1-803-329-9712 - *OUTSIDE THE U.S.*
Fax: 1-803-329-9743

ITS@SENSAFE.COM
WWW.SENSAFE.COM
www.poolcheckonline.com



The purpose of the Ready Snap® 4 Calibration Verification Solution is to verify the working condition of your eXact® photometer and ensure correct testing procedure is followed. The Ready Snap® 4 Calibration Verification Solution is ideal for regulators and inspectors who require a strict verification method. Each ampoule contains a colored liquid which simulates a predetermined analyte value. If your photometer is tested using this Ready Snap® 4 Calibration Verification Solution and the generated results are within the acceptable range, as specified in your eXact® photometer manual, then calibration is verified. Please follow these important instructions.



A

1. Turn on the eXact® Micro Photometer and select the Menu for calibration solution testing (see *Table 1*).



B

2. Fill the CELL to capacity (4mL) with deionized or distilled water. Discard this water sample followed with a quick shake of the meter to empty the CELL of the remaining water drops. Repeat this rinsing process two more times with deionized or distilled water. Finally, refill the CELL to capacity (4mL) with deionized or distilled water.



C

3. Press the ZERO/ON button to zero the meter for the mode selected from *Table 1*. Discard this water sample followed with a quick shake of the meter to empty the CELL of the remaining water drops.



4. Break open one Ready Snap® 4 Blue Calibration Verification Solution in a safe manner with a paper towel as shown (see *Images A and B*). NOTE: Do not break open at poolside or around children or pets. DO NOT CONSUME SOLUTION.

5. Take plastic pipette supplied, squeeze bulb with finger and thumb to expel air, and dip pipette tip to the bottom of ampoule and release squeezed bulb slowly to fill pipette (see *Image C*). Transfer filled pipette liquid to photometer CELL (see *Image D*).

6. Discard this first blue liquid sample followed with a quick shake of the meter to empty the CELL of the remaining colored water drops.

7. Repeat Step 5 and fill the CELL to 4mL capacity with the blue reference solution (see *image D*).

8. Press the READ button to start the 20 second countdown.

9. Verify the display value against the “Assigned Value” located in your eXact® photometer manual. If your value falls within the “Acceptable Value” range, your Photometer is functioning properly.

10. If your value is outside the “Acceptable Value” range, discard the blue solution in the CELL and refill as in Step 5. Then go to Step 8 and retest a fresh blue Calibration Verification Solution sample.

If you continue failing to obtain the acceptable value, retry Step 1 to Step 9 or call (803) 329-9712 for technical support.

Table 1 – Meter MENU Selections for READY SNAP® 4 Solution

eXact® Photometer	MENU selection for blue calibration solution testing
Micro 8	FE2, NH3
Micro 20	TR2 ₇ *
Eco-Check	NH4

R061913

*Subscript indicates the Select Group value for the Micro 20 photometer

Reorder Information

PRODUCT	PART NO.	TESTS	USES
Ready Snap® 1	480901	10	Method Verification Solution for TA, PH, PO4, CA, TH, CH, CY, CU, NO3, MN, AI, CR, F, SO4, PB, and TDS
Ready Snap® 2	480902	10	Method Verification Solution for AS, FE, PO4, NH3, NO3, and MN
Ready Snap® 3	480903	10	Calibration Verification Solution for eXact® Micro 7+, eXact® Micro 10, eXact® Micro 20, and eXact® Eco-Check
Ready Snap® 4	480904	10	Calibration Verification Solution for eXact® Micro 8, eXact® Micro 20, and eXact® Eco-Check
Ready Snap® 5	480905	10	COMING SOON
Ready Snap® 6	480906	10	COMING SOON

Contact Information

For US Inquiries and Re-Orders:

Industrial Test Systems, Inc.

1875 Langston Street, Rock Hill, SC 29730 USA

Phone: 1-800-861-9712 - *INSIDE THE U.S.*

1-803-329-9712 - *OUTSIDE THE U.S.* Fax: 1-803-329-9743



ITS@SENSAFE.COM

www.poolcheckonline.com

WWW.SENSAFE.COM



Industrial Test Systems, Inc.

Innovators of Water Quality Testing

For European & Middle East Inquiries and Re-Orders:

ITS Europe, LTD

The UK Centre for Homeland Security - Building 7, Chilmark

Salisbury, Wiltshire SP3 5DU, United Kingdom

Tel: +44 (0)1722 717911 Fax: +44 (0) 1722 717941

SALES@SENSAFE.COM

www.poolcheckonline.com

WWW.ITSEUROPE.CO.UK



ITS Europe Ltd.

Innovators of Water Quality Testing