

Temperature Compensation Chart

Because the temperature of the tested water will effect the result, you should adjust the *dip time* according the following chart:

°C / °F	Dip Time
0 / 32	20
1 / 34	19
2 / 36	17
3 / 37	15
4 / 39	15
5 / 41	14
6 / 43	14
7 / 45	13
8 / 46	13
9 / 48	12
10 / 50	12
11 / 52	11
12 / 54	11
13 / 55	10
14 / 57	9
15 / 59	9
16 / 60	8
17 / 62	8
18 / 64	7
19 / 66	7
20 / 68	6
21 / 70	6

°C / °F	Dip Time
22 / 72	6
23 / 73	5
24 / 75	5
25 / 77	5
26 / 79	5
27 / 81	5
28 / 82	5
29 / 84	5
30 / 86	5
31 / 88	5
32 / 90	5
33 / 91	5
34 / 93	5
35 / 95	5
36 / 97	5
37 / 99	5
38 / 100	5
39 / 102	5
40 / 104	5

Note: Measure the temperature of the water sample within $\pm 2^{\circ}\text{C}$ / $\pm 4^{\circ}\text{F}$